BRUHAT BANGALORE MAHANAGAR PALIKE

Project Division, Yelahanka Zone

SCHEDULE - B

Grants of Infrastructure up gradation of Yelahanka Govt Hospital

SNo	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
1	3.00	VVIP Excecutive Chair – : Fire retardant breathable mesh for backrest and Polyester fabric for seat. Black colour on seat and back (TW 182 Mesh & 5060 fabric seat) Base : 700 mm diameter raised Nylon base with 60 mm castors for stability and easy movement . Seat cover : Casement cloth protection for seat understructure Adjustable Arms : Moves up or down for shoulder and upper body support. The firm polyurethane arm pads adjust vertically to accommodate different arm lengths from the shoulder to the elbow (TSA012 – value lab pad) Lumbar Support: "S' shaped backrest follow's the natural curve of the back for lower back support. Additional support by static polypropylene lumbar pad on back rest . Seat Contour : Seat pan depth is designed so that the forward most point on the lumbar support is placing gentle pressure on the lower 5 vertebrae on the spine. (2018 – Nuclear Seat) Synchro tilt with single lock : Chair Back reclines synchronously to a 2-to-1 ratio to seat angle. Allows user to recline while keeping seat cushion relatively level to floor. Use the tilt lock for upright posture like keyboarding . Pneumatic height : Change the height of the chair relative to the floor by 100 mm . Headrest : Fixed head rest for neck support	NO	12381.06	Rupees Twelve Thousand Three Hundred Eighty One And Paise Six Only
2	10.00	Doctor's Executive Chair : 1. High back chair with head support. 2. The double mesh back should be designed for comfort and support. 3. Adjustable seat height, head rest height & Hand rest angle 4. Adjust the height using the gas lift to find that comfortable position, and the nylon castors make it easy to move around your space. 5. Five-part Chrome legs 6. Height 48" + 2" 7. Width 26" + 2" 8. Depth 20" + 2" Should be ISI / CE certified	NO	9285.79	Rupees Nine Thousand Two Hundred Eighty Five And Paise Seventy Nine Only
3	6.00	Cardiac Cots: 1.□ Laminated board Top.□ 2.□ Tubular steel frame work.□ 3.□ Mounted on four 5cm. castors.□ 4.□ Epoxy powder coated.□ 5.□ Top size: -106(L)*45(W) cm.Height: -95cm	NO	3713.92	Rupees Three Thousand Seven Hundred Thirteen And Paise Ninety Two Only
4	12.00	Steel Racks: 1. Steel rack made of slotted angles of size 2"x2"x3mm thick, 2. 5 compartments of six 6' x 1 ¼'x 3' length. 3. Paint: Epoxy powder coated in seamen's Gray Colour with minimum 50 microns 4. Should be ISI / CE Certified	NO	6190.53	Rupees Six Thousand One Hundred Ninety And Paise Fifty Three Only
5	10.00	Patient Cots (Plain):1. Size-78"Lx36"Wx24"H, 1 ¼" dia 14 SWG tubular Head and leg bows ERW Tube.2. The head bow and leg bows are 42" and 36" Height respectively. Head and leg bows are 14 gauge, 1" dia ERW cross bar and ½" dia 16-gauge ERW three vertical support bars.3. Top Sheet 18 SWG CRCA Sheet, double Press bent uniformly on all the four sides, 2" vertically and ¾" further inside. A 6" broad x 2" deep full length 16-gauge square cross section rib is reinforced to the top sheet 2 Nos. M.S. Gages using 10 SWG. 4. Zinc plated ½" nuts & bolts, 4 Nos. of Shoes made up of white nylon are provided. 5. Epoxy powder coated with seamen's Gray Colour with 50 microns minimum coating.	NO	6190.53	Rupees Six Thousand One Hundred Ninety And Paise Fifty Three Only

Santhosh.D

Contractor

Executive Engineer
Project Division
Yelahanka Zone

No	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
6	3.00	Patient Cots with side railings: 1. Efficient Height adjustable, manually operated, two section cots with perforated top. 2. The size of approximately 84" length, 36" width and an approximate height adjustable from 20" to 33" with ±5% tolerance 3. The head bow and leg bows are 42" and 36" Height respectively. Head and leg bows are 14 gauge, 1" dia ERW cross bar and ½" dia 16-gauge ERW three vertical support bars. 4. Having a Trendelenburg and reverse Trendelenburg movements. 5. Detachable and collapsible Sturdy railing, made of stainless steel of minimum 25mm diameter and 16-gauge tubes Capsule tubular frame, with a locking facility for raised position.	NO	9905.43	Rupees Nine Thousand Nine Hundred Five And Paise Forty Three Only
7	30.00	Computer Tables and Chairs for Departments excluding e-hospital:Legs : Durable steel powder coated under structure (Broken White) Accessories : Two Wire management holes at either ends of the table to access your cables.Multipurpose & Convenient : table can be kept in your office as well as your home. Its functional design allows you to use this table as a study desk, computer desk, writing table etc. Chair Material : Fire retardant breathable mesh for backrest and Polyester fabric for seat. Black colour on seat and back (TW 182 Mesh & 5060 fabric seat) Base : 700 mm diameter Nylon base with 50 mm castors for stability and easy movement . Seat cover : Injection moulded polypropylene seat edge and understructure protection .Adjustable Arms : Moves up or down for shoulder and upper body support. The firm polyurethane arm pads adjust vertically to accommodate different arm lengths from the shoulder to the elbow (TSA012 – value lab pad) Seat Contour : Seat pan depth is designed so that the forward most point on the lumbar support is placing gentle pressure on the lower 5 vertebrae on the spine. (2018 – Nuclear Seat)	NO	22586.00	Rupees Twenty Two Thousand Five Hundred Eighty Six Only
8	30.00	Patient's Waiting Chairs (5 Seater, Wooden):5 Seater Stainless Steel Waiting Chair	NO	8667.13	Rupees Eight Thousand Six Hundred Sixty Seven And Paise Thirteen Only
9	20.00	Patient's waiting Chairs (3 Seater, Wooden):Stainless Steel Silver Three Seater Waiting Chair, for Hospital,Office etc., Seating Capacity: 3	NO	5571.87	Rupees Five Thousand Five Hundred Seventy One And Paise Eighty Seven Only
10	3.00	Drugs Transport Trolleys:1. Trolley with 25 mm diameter SS tubular frame2. Drawers 6 number possible of adequate size with top flat stainless steel3. Size – 36-inch L x 20-inch W x 60-inch H.4. 1 rows of hand out bins of different size & colour to hold different sizes5. Light weight plastic box with drawers of different sizes and colours to hold emergency drugs6. Facility to carry monitor and Defibrillator with Stainless steel saline rod-one.7. Nylon Castor wheels of 12.5 cm dia-Two having locking arrangement.	NO	22286.49	Rupees Twenty Two Thousand Two Hundred Eighty Six And Paise Forty Nine Only
11	2.00	3 Bucket System (Mopping):Material Stainless Steel Applications Mopping system Bucket Capacity 30-60 L, 48 Lt. Trolley Material Stainless Steel Model APC-119/3SS Number of Buckets 3 Usage/Application Hospitals Brand Amsse Color Red Down Press Wringer yes Minimum Order Quantity 1 Set	NO	13619.36	Rupees Thirteen Thousand Six Hundred Nineteen And Paise Thirty Six Only

Santhosh.D

Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
12	20.00	Fibre Pallets (Pharmacy) MainStore :Fiber Pallets 4"x 5"	NO	6190.53	Rupees Six Thousand One Hundred Ninety And Paise Fifty Three Only
13	3.00	Dispensing Chairs:Material: Fire retardant breathable mesh for backrest and Polyester fabric for seat. Black colour on seat and back (TW 182 Mesh & 5060 fabric seat) Base : 700 mm diameter raised Nylon base with 60 mm castors for stability and easy movement . Seat cover : Casement cloth protection for seat understructure Adjustable Arms : Moves up or down for shoulder and upper body support. The firm polyurethane arm pads adjust vertically to accommodate different arm lengths from the shoulder to the elbow (TSA012 – value lab pad) Lumbar Support: "S' shaped backrest follow's the natural curve of the back for lower back support. Additional support by static polypropylene lumbar pad on back rest . Seat Contour : Seat pan depth is designed so that the forward most point on the lumbar support is placing gentle pressure on the lower 5 vertebrae on the spine. (2018 – Nuclear Seat) Synchro tilt with single lock : Chair Back reclines synchronously to a 2-to-1 ratio to seat angle. Allows user to recline while keeping seat cushion relatively level to floor. Use the tilt lock for upright posture like keyboarding .Pneumatic height : Change the height of the chair relative to the floor by 100 mm .	NO	12381.06	Rupees Twelve Thousand Three Hundred Eighty One And Paise Six Only
14	20.00	Cushion Chair with Arm:Brand Maruthi Color option available Seat Material Fabric Size standard Design Type Customized No Of Wheels 5 Revolving Chair Yes Material cushion Condition New Minimum Order Quantity 1 Piece	NO	3713.92	Rupees Three Thousand Seven Hundred Thirteen And Paise Ninety Two Only
15	2.00	Blood Drawing Chair:Color Blue Material Stainless steel General Use Blood banks Brand Agile Number Of Seats 1 Weight 120 kg Automation Grade Remote Controlled Usage/Application Hospitals,Blood Banks Voltage 230 V	NO	80478.83	Rupees Eighty Thousand Four Hundred Seventy Eight And Paise Eighty Three Only
16	30.00	Wooden Stool:Surface Finish Polished Material Wooden Shape Square Feature Termite Free	NO	2476.60	Rupees Two Thousand Four Hundred Seventy Six And Paise Sixty Only
17	100.00	Plastic Chair:��? 1.????????? Should be made from the finest polypropylene.???????????????? 2.????????? Should be Tough and lightweight.???????????????? 3.????????? Should be Good quality and Modern designed.???????????????? 4.????????? Should be Weatherproof and virtually maintenance free.	NO	867.11	Rupees Eight Hundred Sixty Seven And Paise Eleven Only
18	20.00	Wooden Tables with Drawers:Surface Finish PolishedMaterial Particleboard Size 4x2 Feet Shape Rectangular Warranty 1 Year Seating Capacity 1 Seater	NO	7181.37	Rupees Seven Thousand One Hundred Eighty One And Paise Thirty Seven Only

Santhosh. D

Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
[illegible]	[illegible].00	Wooden Almaraha With Glass Doors casuality to keep registers	NO	24763.09	Rupees Twenty Four Thousand Seven Hundred Sixty Three And Paise Nine Only
20	5.00	Sterile Equipments /bins Storage Trolly:Capacity 300 kg Brand Jindal Material Mild Steel	NO	5571.87	Rupees Five Thousand Five Hundred Seventy One And Paise Eighty Seven Only
21	10.00	Folding Cots:Material: Steel Pipe Capacity: 120 kg	NO	9905.43	Rupees Nine Thousand Nine Hundred Five And Paise Forty Three Only
22	5.00	Almariha Steel:1. Size: 78" x 36" x 19" with 3 Shelves to be fixed having 4 compartments, 2. 1 compartment with Locker having lever lock placed in a suitable place. The locking system must be mounted securely to give robustness and long life. 3. Double door with three way locking device with duplicate keys, made out of CRCA Sheet of 18 gauge with de rusted synthetic enamel Siemens Gray colour paint. 4. The edges of doors and shelves should be properly bent to U Shape neatly and finished.	NO	13619.36	Rupees Thirteen Thousand Six Hundred Nineteen And Paise Thirty Six Only
23	6.00	Hospital Surrondings Dust Bins:RC Ventures Plastic Wheeled Dust Bin for Outdoor Usage/Application Outdoor Brand RC Ventures Material Plastic Capacity 120 Liters Number Of Wheels 2	NO	3713.92	Rupees Three Thousand Seven Hundred Thirteen And Paise Ninety Two Only
24	5.00	Wheel Barrows:Model Name/Number RJTE 005 Load Capacity 300 ltrs	NO	30953.62	Rupees Thirty Thousand Nine Hundred Fifty Three And Paise Sixty Two Only
25	5.00	Stretcher: 1. Overall Size: Length 84". Width 24". Height 32" 2. Design Removable stretcher top made up of S.S. 20 SWG 304 Grade with full polish with removable cushion made up of 2" foam with inner cloth cover and S.T. quality Green Drill back rexine. 3. The top should be dished to a depth of not less than 1" not more than 2", further fixed on 1" diameter 18 SWG S.S. tubular frames. 4 legs. of 1" diameter 18 SWG supporting floor studs also to be provided, at the bottom of stretcher with nylon shoes. 1.1/4" Sq. 16 SWG ER W pipes used for vertical and horizontal supporting frame of stretcher. Stretcher handle S.S. 18 SWG 304 grade round shape with full polish. 6" diameter 4 nos. swivelling castor rubber wheels 4. Should have suitable stand for B-Type oxygen cylinder. 5. Epoxy powder coating in seamen's grey colour with minimum 50 microns. 6. Should be ISI / CE Certified.	NO	14857.66	Rupees Fourteen Thousand Eight Hundred Fifty Seven And Paise Sixty Six Only
26	5.00	Wheel Chair:□□□□□ 1.□□□□□□□□ Floor to seat height:650 mm to 720 mm.□□□□□□□□□□□□□□□□□□ 2.□□□□□□□□ Overall Size Width:1000 mm to 1100 mm. Length 910 mm to 950 mm. Height 480 mm to 510 mm.□□□□□□□□□□□□□□□□□□ 3.□□□□□□□□ Design: 1" Diameter 16 SWG ERW tubular frame fitted with metal seat and back. Adjustable horizontal aluminium foot rest 90□ - 120□ with tread flats. Half round Wooden Hand rest with 1" cushion. Solid rubber tyre wheels 24" x 1 W' fitted with self-propelling hoops and brake arrangement. Full length axle should be provided to fix the rickshaw wheel.	NO	6190.53	Rupees Six Thousand One Hundred Ninety And Paise Fifty Three Only

Santhosh D

Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
27	13.00	Cardiac Monitors Adult (10) /Paediatric(3): 1. Should have the facility of monitoring ECG, RR, Sp02, NIBP and Dual Temp, for Adult, Pediatric &Neonatal applications. 2. Should have integrated colour display of atleast 12' or more LCD touch screen display. 3. Should have non-volatile Graphical & Tabular trend 4. Should operate independently on both mains and battery. Battery backup for at least 60 Minutes built in battery which is rechargeable.facility for at least 30 hrs. 5. Should have alarm limits with alarm levels and alarm indication (visual as well as audio). 6. 3/5 lead ECG measurement, Respiration and simultaneous monitoring of two temperatures 7. Unit should be supplied with following accessories: 8. 3 lead ECG cable w 12. Pulse Oximeter probe: Reusable Adult and Pediatrics. – Two eachith disposable electrodes. 9. 100 no of disposable electrodes 10. NIBP CUFF – Adult & Pediatric- Two sets each 11. Temp probe Tape on skin 13. Monitor should have built in Electro Surgical Unit &	NO	76386.83	Rupees Seventy Six Thousand Three Hundred Eighty Six And Paise Eighty Three Only
28	20.00	Finger Pulse Oxymeter Adult /New Born:a. SpO2 measurement range at least 40-70 and 70 to 99 %, minimum gradation 1%.b. Accuracy of SpO2 better than} 1% for range 40-70 and better than} 3% for range 70-99.c. Pulse rate range at least 30 to 240 bpm, minimum gradation 1 bpm.d. Accuracy of pulse rate better than 5 bpm.e. Signal strength or quality to be visually displayed.f. Audio-visual alarms required: high and low SpO2 and pulse rate (operator variable settings), sensor disconnected, sensor failure, low battery.g. TFT Screen.h. Plethysmograph (may be in form of bar) display is mandatory.2. Settings should have minimum 24 hrs trend memory for SpO2 & PR.3. User's interface Easily accessible touch button to operate the machine.	NO	25854.10	Rupees Twenty Five Thousand Eight Hundred Fifty Four And Paise Ten Only
29	2.00	Double Surface Phototherapy Unit (Laser): 1. Phototherapy should be based on advanced LED technology 2. It should have sufficient LED so that the irradiance should cover the entire treatment area. 3. Should have a height adjustable mechanism, treatment distance to the range of 25 to 45 cms should be possible. 4. Wave length of LED should be in the range of 420 –470 nm and irradiance level should be higher than normal blue tube lights. 5. The unit should provide a minimum of irradiance 10Watts/m2. 6. Lamp source should be continuous tiltable to ±90-degree angle to cover the entire treatment area. 7. System should be height adjustable with built-in non resettable timer 8. Should work with input 200 to 240Vac 50 Hz supply. 9. Should have safety certificate from a competent authority CE / FDA (US)/ISI	NO	76386.83	Rupees Seventy Six Thousand Three Hundred Eighty Six And Paise Eighty Three Only
30	9.00	Electric Suction Apparatus, 1HP (3) /1/2 (6)HP new Born: 1. Should be a complete stainless-steel body / any non-corrosive material 2. Should have pump air flow rate minimum 30 – 40/min. 3. Should have vacuum above – 500 mm hg4. Should have three step vacuum adjustments – high, medium and low. 5. Should have secretion container of 2ltr capacity 6. The hose connection between the suction system and secretion jar should be less than 15 cms 7. Should have hygienic bacterial filter with overflow protection 8. Should have sound level maximum of 55 to 65 db 9. Should have motor capacity of minimum of 0.5 HP (1/2 HP) working power of 220/230 V single phase. 10. Should have power supply indicator, suction on indicator, Unit on/off 5amps DP switch, protector fuse and minimum chord length of 5M with 5 Amps 3 pin top. 11. Should be mounted on easy moving castor wheels.	NO	17039.66	Rupees Seventeen Thousand Thirty Nine And Paise Sixty Six Only

Santhosh. D

Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
31	2.00	Electrical Cautery Machine:���Monopolar cut: max 400W. Low, Pure, Blend & Endo ii. Monopolar coag: 120W. Desiccate, Fulgurate & Spray iii. Bipolar coag: 120W. Macro, Standard and Precise iv. Bipolar Cut: 100W v. Vessel Seal: L-5, 150W, Auto-stop & regrasp indicators vi. Programmable power settings: 10 programs vii. Tissue sensing technology viii. Patient plate contact quality monitoring ix. HF leakage monitoring Accessories * Silicon Split Patient Plate * Cable for Patient Plate *Disposable handswitch, 3 Pin *Electrode set *Bipolar forcep * Bipolar cable *5mm laparoscopic vessel ealing instrument with cable *Open biclamp with cable *Three-pedal Footswitch	NO	411312.66	Rupees Four Lakh(s) Eleven Thousand Three Hundred Twelve And Paise Sixty Six Only
32	2.00	Battery Operated Power Drill (P-Comp with Respac Software)	NO	176276.86	Rupees One Lakh(s) Seventy Six Thousand Two Hundred Seventy Six And Paise Eighty Six Only
33	1.00	ENT Microscope (Optive):• Manual 5 Step Magnification changer with apochromatic optics. Magnification Factor 0.4x,0.6x,1.0x,1.6x,2.5x • Objective Lenses f-250mm or 300mm with manual focusing mechanism • Objective Lens Fixed working distance f=400mm • Straight Binacular tube f=170mm interpupillary distance adjustment from 55mm to 75mm • Pair of wide field push –in eyepieces 10x with sleeves and magnetic locks dioptre setting from 6D to -6D • 120 Degree Coupling • Front Hand Grip • Coaxial Fiber Optic Illumination System with LED Illumination. • Stable and study floor stand on four lockable castors, column Spring Balance articulated arm, Carrier arm Power supply , unit Light Guide ,2m power cable .	NO	763866.38	Rupees Seven Lakh(s) Sixty Three Thousand Eight Hundred Sixty Six And Paise Thirty Eight Only
34	1.00	Otoscope Welch Allyn:���������� 1.��������� Should be a convenient pocket type otoscope.���������� 2.��������� Should be provided with a halogen light source.���������� 3.��������� Should be able to detach the otoscope head.���������� 4.��������� Should provide no reflections and obstructions.���������� 5.��������� Should provide detachable accessories of various sizes.���������� 6.��������� Should have in built rechargeable battery. Recharge should be possible with direct mains supply	NO	277342.33	Rupees Two Lakh(s) Seventy Seven Thousand Three Hundred Forty Two And Paise Thirty Three Only
35	12.00	Oxygen Flow meter with Regulator:�� 1.��������� Internal reseating relief valve protects against over-pressurization���������� 2.��������� Sintered filter for additional safety and to extend regulator life���������� 3.��������� 2in Diameter gauge���������� 4.��������� 15LPM Oxygen models provide 85 to 90LPM flood maximum���������� 5.��������� 7LPM models restricted to 10LPM maximum	NO	2938.14	Rupees Two Thousand Nine Hundred Thirty Eight And Paise Fourteen Only

Santhosh.D

Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
36	6.00	ECG MACHINE :��? 1.????????? Simultaneous 12 Channel ECG recording with 12 lead simultaneous acquisition????????? 2.????????? Should have visual alarm for open lead????????? 3.????????? Should have a digital display of 12 channel ECG????????? 4.????????? QWERTY Alphanumeric keyboard????????? 5.????????? Built-in ECG Parameters measurements and Interpretation????????? 6.????????? Minimum 40 ECG Storage inbuilt memory.????????? 7.????????? 3 Operating modes: Automatic, Manual and Rhythm????????? 8.????????? Should have a maintenance free digital thermal array printer????????? 9.????????? Printer should work with standard thermal paper (should be available in Local Market)???????? 10.???????? Should have 12 lead ECG preview display before taking printouts and should have printer on/off selection.???????? 11.???????? Should have ECG lead annotation facility???????? 12.???????? Machine should have sufficient battery backup for taking at least 25 nos ECG on a fully charged battery ???????? 13.???????? Should supplied with 2 patient cable sets, 8 clips on electrodes, 12 chest electrodes with silicone rubber bulb, 12 packets of recording paper,1 bottle of jelly and 12 nos. reusable button type electrode???????? 14.???????? Should operate on	NO	149264.00	Rupees One Lakh(s) Forty Nine Thousand Two Hundred Sixty Four Only
37	4.00	Debrilator Adult (3) And Peadatric (1): 1.????????? Technical characteristics (specific to this type of device):a.?????? Unit should be lightweight compact and portable.b.????? Unit should have facility for Automatic External Defibrillation and manual defibrillation.c.?????? Should be able to deliver shock from 50-200 joules in biphasic mode via metal chest pads.d.????? Should having design protection to avoid passage of current to the user.e.????? The whole system should have an inbuilt recorder; TELEMETRY NOTRECOMMENDED.????????? 2.????????? Settings: Manual AND Automatic.????????? 3.????????? User's interface: The monitor should have a TFT colour display with a three-channel display.????????? 4.????????? Software and/or standard of communication (where ever required): Inbuilt????????? 5.????????? adjustable heart rate alarm as well as paddles & ECG cable disconnection alarms.????????? 6.????????? Power Requirements: 220 to 240V, 50 Hz.????????? 7.????????? Battery operated: Rechargeable battery backup of approximately 5 hours.????????? 8.????????? Protection: Electrical protection by resettable over current breakers or replaceable fuses, Fitted in both live and neutral	NO	309534.26	Rupees Three Lakh(s) Nine Thousand Five Hundred Thirty Four And Paise Twenty Six Only
38	1.00	Keratometer:Specification Measurement Ranges Corneal radius 6.4 - 9.4 mm in 0.05 mm step Corneal power 52- 36 D diopter in 0.25 D step Corneal astigmatism axis 00 - 1800 Electrical Specifications Light source LED Bulb Power supply ~110-230 V, 50-60 Hz Physical Parameters Dimension (L x B x H) 19 x 7 x 16 inches Weight Approx 7 Kg	NO	58758.95	Rupees Fifty Eight Thousand Seven Hundred Fifty Eight And Paise Ninety Five Only

Santhosh. D

Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Devarajappa	
				In figures	In words
39	1.00	FIVE STEP SLIT LAMP: Stereo Microscope Type Galilean Magnification Changer Drum rotation Working Distance 100 mm Objective Lens Focal Length f 100 mm Eye Pieces (Large Optics) 12.5x Total Magnifications 6x, 10x, 16x, 25x, 40x Real Field of View 43, 27,16,11, 7 mm Inter Pupillary Distance 55 - 75 mm Diopter Adjustment Range -6 D to +6 D Common Specifications Slit image Width 0 to 14 mm continuously variable Slit image Height 0 to 14 mm continuously variable Illumination Field Diameter in mm 0.2, 2, 3, 4, 6, 8 & 14 mm Slit Rotation 0° to 180° Slit Inclination 5°, 10°, 15°, 20° Sclerotic Scatter Knob for Indirect illumination & Retro illumination Filters Heat absorbing, Green (Red free), Cobalt blue Illumination 12 V- 30 W Halogen lamp / 3.4 V-700 mAh LED Operation 3 Dimensional single handed operation Instrument base movement 99 mm (in/out), 30 mm (up/down),118 mm (left/right) Physical Parameters Dimensions L x W x H) 725 x 460 x 555 mm Weight net 20 Kg and gross 27 Kg (approx) Chin rest movement range 120 mm Electrical Specifications Power supply ~110-230 V, 50-60 Hz Power consumption 45 W max(Halogen) / 2.4 W max(LED) Bulb Led Bulb	NO	225860.00	Rupees Two Lakh(s) Twenty Five Thousand Eight Hundred Sixty Only
40	1.00	FLASH AUTO CLAVE:Size -Overall W = 40 to 50cm L = 40 to 50cm H = 30 to 40cm approx -Chamber 20cm x 35cm L approx. Electrical: -Supply 220-240V Construction Stainless steel chamber, door and trays Instrumentation Cycle status display. Digital temperature display, Pressure guage, Safety Pressure safety valve, Door safety catch, Overheat Protection with warning indicator, Chamber pressure indicator, Pressure door lock. Sterilizing: -Sterilizing time: At 134 deg.C---------5 to 12 min; At 121 deg.C--------15 to 25 min. Drying time approx 15minutes when selected	NO	433347.76	Rupees Four Lakh(s) Thirty Three Thousand Three Hundred Forty Seven And Paise Seventy Six Only
41	1.00	TENS (BOTH FARADIC & GALVANIC):Advanced Programmed T..E.. N.. S.. Therapy Unit (Dual Channel)LCD Display12 preset ProgrameBurst100HZBurst 200 HzConstant 2HzConstant 10 HzConstant 20 HzConstant 40 HzConstant 80 HZConstant 100 HzConstant 120 HzConstant 150 HzSurge 100HzSurge 200 HzPower supply : 90V to 270V (SMPS Power supply)Unit will be supplied with all standard accessories :Main Unit : 01 No.Electrodes : 04 Nos.Main cable : 01 No.Wire Set : 02 Nos. (4 output)Power Socketwill be provided free of cost as per country requirement.	NO	23568.00	Rupees Twenty Three Thousand Five Hundred Sixty Eight Only

Executive Engineer
Project Division
Yelahanka Zone

Contractor

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
42	2.00	NST Machine with probe :�� 1.��������� It should have Numerical display of FHR, TOCO & Foetal Movements Counts��������� 2.��������� It should have min of 15cm Display for displaying Wave forms of FHR & TOCO��������� 3.��������� It should have Automatic Foetal Movement Detection��������� 4.��������� It should have minimum of 2 Hours Battery Backup��������� 5.��������� It should be provided with Computer Foetal monitoring Software and Data cable for storage & Reload��������� 6.��������� It should be easy to use & take print out��������� 7.��������� It should have inbuilt Thermal recorder��������� 8.��������� It should have inbuilt memory of 10 Hrs Graphical Recording��������� 9.��������� It should be provided with Aluminium cart with wheels for easy Movement of the machine******** 10.******** It should have built in Network Capability******** 11.******** Should be BIS / CE / FDA certified product.	NO	127660.00	Rupees One Lakh(s) Twenty Seven Thousand Six Hundred Sixty Only
43	2.00	Foetal Doppler :��������� 1.��������� Alarm alert to indicate baby cardiac and tachycardia.��������� 2.��������� Dry nickel cadmium battery with charger.��������� 3.��������� Spiral cable for the probe with length of 1 mtr. Life for battery of at least 6 to 8 hours.��������� 4.��������� Volume mode adjustable.��������� 5.��������� Pouch for the machine.��������� 6.��������� Digital display of FHR.��������� 7.��������� Should be ISI/CE/FDA certified	NO	9820.00	Rupees Nine Thousand Eight Hundred Twenty Only
44	10.00	Infusion Pump :1.******* Should be operated on drip rate Peristaltic finger pump method.2.******* Should compatible with most of the IV set (macro/micro drip sets).3.******* Should have the following flow rates.4.******* IV Set ml/hr drops/min15 drops/ml 3~450ml/hr 1~100drops/min20drops/ml 3~450ml/hr 1~100drops/min60drops/ml 1~100ml/hr 1~100drops/min5.******* Should have a flow rate accuracy of ±10% and drip rate accuracy of ±2%.6.******* Should have a volume infused display from 0 to 999.9ml.7.******* Should have a purge andKVO facility.8.******* Should have an audible and visual alarm for occlusion pressure, air alarm, door open,empty, low battery.9.******* Should have a LCD display with backlight and graphical display of infusion Shouldhave a minimum 2hr battery back up at highest delivery rate.10.** Should work with input 200 to 240Vac 50 Hz supply.11.** Should have safety certificate from a competent authority CE / FDA (US) / ISI.	NO	56956.00	Rupees Fifty Six Thousand Nine Hundred Fifty Six Only
45	5.00	Geyser:V-guard geyser 25 litre price 5 star	NO	13748.00	Rupees Thirteen Thousand Seven Hundred Forty Eight Only
46	2.00	Double Door Refrigerator 810 Litres:Samsung refrigerator double door 700 ltr	NO	80524.00	Rupees Eighty Thousand Five Hundred Twenty Four Only
47	2.00	Single Door Refrigerator 165 Litres:LG refrigerator single door 180 ltr	NO	14730.00	Rupees Fourteen Thousand Seven Hundred Thirty Only
48	4.00	Single Door Mini Refrigerator Litres:LG refrigerator mini	NO	14730.00	Rupees Fourteen Thousand Seven Hundred Thirty Only

Santhosh. D

Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
	10.00	Split A.C :INSTALLATION CHARGES+Hitachi 1.5tr tr Split Ac 5-Star	NO	58774.66	Rupees Fifty Eight Thousand Seven Hundred Seventy Four And Paise Sixty Six Only
50	2.00	Microave :LG 28 L Convection Microwave Oven	NO	17676.00	Rupees Seventeen Thousand Six Hundred Seventy Six Only
51	3.00	Stove (Electrical):Electrical stove	NO	12381.06	Rupees Twelve Thousand Three Hundred Eighty One And Paise Six Only
52	2.00	Colour Foetal Doppler:Foetal Doppler 1. Alarm alert to indicate baby cardiac and tachycardia. 2. Dry nickel cadmium battery with charger. 3. Spiral cable for the probe with length of 1 mtr. Life for battery of at least 6 to 8 hours. 4. Volume mode adjustable. 5. Pouch for the machine. 6. Digital display of FHR. 7. Should be ISI/CE/FDA certified	NO	9905.43	Rupees Nine Thousand Nine Hundred Five And Paise Forty Three Only
53	20.00	Autoclave Steel Bins 24"x24":Stainless Steel 304G	NO	5876.29	Rupees Five Thousand Eight Hundred Seventy Six And Paise Twenty Nine Only
54	10.00	Autoclave Steel Bins 18"x18"	NO	5288.07	Rupees Five Thousand Two Hundred Eighty Eight And Paise Seven Only
55	10.00	Autoclave Steel Bins 15"x15"	NO	3525.38	Rupees Three Thousand Five Hundred Twenty Five And Paise Thirty Eight Only
56	60.00	Bedside Lockers:Bed Side Locker 1. Size: H32" x W16" x D16" 2. Size of the Cupboard: H16" x W14"x D16" located in the centre 3. Cupboard made of CRCA steel sheet with double hinged door and locker facility 4. Locker top to have S.S. Sheet super imposed with raised edges on three sides 5. Framework of 1" square M.S. and legs with rubber feet. Full frame and cupboard to be powder coated in Seamen's Grey colour with 50 microns coating. 6. Should be ISI / CE Certified	NO	2476.60	Rupees Two Thousand Four Hundred Seventy Six And Paise Sixty Only
57	10.00	Stainless Steel Foot Step:Stainless Steel Foot Step 1. Overall size: 18"X9"X20" 2. TOP SHEET: 18 gauge S.S. 3. DESIGN: Top fitted with aluminium tread flats. Height 9" and 20" 1" Sq.16 SWG SS Pipes to be used. 4 Nos. Nylon feet 4. Should be ISI / CE Certified	NO	1577.09	Rupees One Thousand Five Hundred Seventy Seven And Paise Nine Only

Santhosh. D

Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
58	10.00	Powder Coated Back Rest:Size 78Lx36Wx24H, Back Rest by Screw rod Mechanism, Powder coated Uniformly perforated Two Sectional CRCA sheet top, Frame work made of Rectangular M.S. Tube, ABS Bow, Collaociable side Railings The above product should conform to applicable ISI Standards in respect of materials, welding process, tests and performance, Finish: Pretreated and Epoxy Powder Coated to be supplied along with Matterss 3' coir n 1" foam	NO	18571.58	Rupees Eighteen Thousand Five Hundred Seventy One And Paise Fifty Eight Only
59	10.00	Foot End Elevators:Size 78Lx36Wx24H, Back Rest & Knee rest by Screw rod Mechanism, Powder coated Uniformly perforated four sectional CRCA sheet top, Frame work made of Rectangular M.S. Tube, ABS Bow, Collaociable side Railings The above product should conform to applicable ISI Standards in respect of materials, welding process, tests and performance, Finish: Pretreated and Epoxy Powder Coated to be supplied along with Matterss 3' coir n 1" foam	NO	47627.00	Rupees Forty Seven Thousand Six Hundred Twenty Seven Only
60	20.00	Plastic Bed Pans	NO	294.60	Rupees Two Hundred Ninety Four And Paise Sixty Only
61	20.00	Plastic Urine Can	NO	117.84	Rupees One Hundred Seventeen And Paise Eighty Four Only
62	20.00	Needle Holder 8" :Length: 8'' b. Tip Configuration: Blunt c. Instruments Type: Needle Holder d. Curvature: Straight e. Working Surface Style: Cross-Serrated with Groove f. Handle: Finger Rings g. Should be high Quality Surgical steel Instruments and Manufactured from medical grade surgical stainless steel h. Should be Sterilizable	NO	589.20	Rupees Five Hundred Eighty Nine And Paise Twenty Only
63	50.00	Needle Holder 6" ,:Length: 6'' b. Tip Configuration: Blunt c. Instruments Type: Needle Holder d. Curvature: Straight e. Working Surface Style: Cross-Serrated with Groove f. Handle: Finger Rings g. Should be high Quality Surgical steel Instruments and Manufactured from medical grade surgical stainless steel h. Should be Sterilizable	NO	491.00	Rupees Four Hundred Ninety One Only
64	250.00	Allis Forceps 4":Length: 4'' b. Instrument Type: Allis Forceps 4" c. Curvature: Straight d. Handle: Finger Rings e. Should be High Quality Surgical Steel Instruments and Manufactured from medical grade Surgical Stainless steel f. Should be Sterilizable	NO	392.80	Rupees Three Hundred Ninety Two And Paise Eighty Only
65	50.00	Cockers St 7":Length: 7'' b. Instrument Type: Cockers St 7" c. Curvature: Straight d. Handle: Finger Rings e. Should be High Quality Surgical Steel Instruments and Manufactured from medical grade Surgical Stainless steel f. Should be Sterilizable	NO	687.40	Rupees Six Hundred Eighty Seven And Paise Forty Only

Santhosh. D

Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
	50.00	Cockers CURVED 7":Length: 7" b. Instrument Type: Cockers St 7" c. Curvature: Curved d. Handle: Finger Rings e. Should be High Quality Surgical Steel Instruments and Manufactured from medical grade Surgical Stainless steel f. Should be Sterilizable	NO	687.40	Rupees Six Hundred Eighty Seven And Paise Forty Only
67	20.00	Scissors Fine Tip 7":Size: 7" b. Curvature: fine tip c. Fully graduated in centimetres d. Should be high Quality Surgical steel instrument and Manufactured from medical grade surgical stainless steel e. Should be Sterilizable	NO	736.50	Rupees Seven Hundred Thirty Six And Paise Fifty Only
68	30.00	Scissors Straight 7":Size: 7" b. Curvature: Straight c. Fully graduated in centimetres d. Should be high Quality Surgical steel instrument and Manufactured from medical grade surgical stainless steel e. Should be Sterilizable	NO	736.50	Rupees Seven Hundred Thirty Six And Paise Fifty Only
69	25.00	Scissiors 6" curved:Size: 6" b. Curvature: Curved c. Fully graduated in centimetres d. Should be high Quality Surgical steel instrument and Manufactured from medical grade surgical stainless steel e. Should be Sterilizable	NO	687.40	Rupees Six Hundred Eighty Seven And Paise Forty Only
70	20.00	Scissors Fine Tip 5":Size: 5" b. Curvature: fine tip c. Fully graduated in centimetres d. Should be high Quality Surgical steel instrument and Manufactured from medical grade surgical stainless steel e. Should be Sterilizable	NO	687.40	Rupees Six Hundred Eighty Seven And Paise Forty Only
71	3.00	Hadson's Tooth Forceps :Length: 4.75" (12cm) b. Handle: Flat, Serrated Thumb c. Curvature: Straight d. Jaw Style: Fine e. Tip Configuration: 1x2 Teeth, Side-Grasping f. Instrument Type: Tissue Forceps g. Material: Surgical Stainless Steel h. Usage: Reusable, Non-Sterile	NO	343.70	Rupees Three Hundred Forty Three And Paise Seventy Only
72	3.00	Hadson's non Tooth Foeceps :Length: 4.75" (12cm) b. Handle: Flat, Serrated Thumb c. Curvature: Straight d. Jaw Style: Fine e. Instrument Type: Tissue Forceps f. Material: Surgical Stainless Steel g. Usage: Reusable, Non-Sterile	NO	343.70	Rupees Three Hundred Forty Three And Paise Seventy Only
73	2.00	Periosteum Elevator Size 5 & 6 each size (1 +1):Length: 5" & 6" b. Handle: Flat and Smooth c. Curvature: Straight d. Jaw Style: Fine e. Tip Configuration: CVD Tip f. Instrument Type: Surgical Elevator g. Material: Surgical Stainless Steel h. Usage: Reusable, Non-Sterile	NO	1374.80	Rupees One Thousand Three Hundred Seventy Four And Paise Eighty Only
74	20.00	Allies Small 3":Length: 3" b. Instrument Type: Tissue Grasping Forceps c. Curvature: Straight d. Handle: Finger Rings e. Should be High Quality Surgical Steel Instruments and Manufactured from medical grade Surgical Stainless steel f. Should be Sterilizable	NO	294.60	Rupees Two Hundred Ninety Four And Paise Sixty Only

Executive Engineer
Project Division
Yelahanka Zone

Santhosh D
Contractor

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
	60.00	Speculum Medium :Shape: Should be Double Ended b. Size: Should be medium in size (blade one size 75◆30mm, other 80◆35 mm) c. Material: Should be high Quality Surgical steel instrument and manufactured from medical grade surgical Stainless steel d. Safety: Protects Patients against Cross infection risks. e. Use: Should be reusable	NO	1335.52	Rupees One Thousand Three Hundred Thirty Five And Paise Fifty Two Only
85	13.00	Anterior Vaginal wall Retractor:Made of 410 Grade Stainless Steel b. Total Length – 25.5CM c. Rings should be oval shaped d. Both sides serrated rings e. Ring internal diameter – 15MM & 18MM f. Matt/ Dull Finish	NO	1119.48	Rupees One Thousand One Hundred Nineteen And Paise Forty Eight Only
86	1.00	Tubual Ligation Set	NO	8347.00	Rupees Eight Thousand Three Hundred Forty Seven Only
87	1.00	Surgical D &C Set	NO	9820.00	Rupees Nine Thousand Eight Hundred Twenty Only
88	30.00	B.P.Handle big:Forgings from Stainless Steel. b. Silk Matte Satin Finish. c. To Hold Scalpel Blades. d. Used Mostly In All Surgery.	NO	294.60	Rupees Two Hundred Ninety Four And Paise Sixty Only
89	10.00	B.P. Handle small:Forgings from Stainless Steel. b. Silk Matte Satin Finish. c. To Hold Scalpel Blades. d. Used Mostly In All Surgery.	NO	235.68	Rupees Two Hundred Thirty Five And Paise Sixty Eight Only
90	2.00	Varicose Veins Stripper	NO	11293.00	Rupees Eleven Thousand Two Hundred Ninety Three Only
91	2.00	Ovum Foeceps (gynec):Size: 9-1/2" (24.1 cm), Straight, Serrated Jaws b. Fully graduated in centimetres c. Should be high quality surgical steel Instrument and manufactured from medical grade Surgical stainless steel d. Should be Sterilizable	NO	1335.52	Rupees One Thousand Three Hundred Thirty Five And Paise Fifty Two Only
92	20.00	Cord Cutting Scissors:Should be very wide blades with large, rounded, blunt tips that slide past each other b. The blades are attached to handles with the proximal end conformed as a ring to facilitate holding and operating with the fingers; both parts are joined by a swivel pin in the centre. c. Should be high quality surgical steel Instrument and manufactured from medical grade Surgical stainless steel d. Should be Sterilizable	NO	608.84	Rupees Six Hundred Eight And Paise Eighty Four Only
93	25.00	Episcotmy Scissiors:Made of 420 grade stainless steel b. ISI/CE/FDA certified & anti rust instrument c. Total length – 14 cm d. Blade length – 4.6 cm e. Blades should be angled f. Rings should be smooth g. Matt/ dull finish	NO	677.58	Rupees Six Hundred Seventy Seven And Paise Fifty Eight Only
94	2.00	Cervical Biopsy Set	NO	2847.80	Rupees Two Thousand Eight Hundred Forty Seven And Paise Eighty Only
95	60.00	Steel Trays medium :Instrument Tray with cover – Stainless steel of 304 grade b. Size: 300x 200 x 65mm	NO	569.56	Rupees Five Hundred Sixty Nine And Paise Fifty Six Only

Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
	20.00	Cheatle Forceps:Made of 410 grade stainless steel b. ISI/CE/FDA certified & anti rust instrument c. Total length – 25.5 cm d. Jaw length – 7.8 cm e. Jaws should be smooth without serration f. Rings should be smooth g. Matt/ dull finish	NO	844.52	Rupees Eight Hundred Forty Four And Paise Fifty Two Only
97	10.00	Suction Tips Different Size:Tubing Material: Medical-grade Polyvinyl Chloride (PVC) Plastic b. tubing size 1/4", length 1.80m, Yankauer handle with control vent and crown tip c. tubing size 1/4", length 3.60m, Yankauer handle with control vent and crown tip d. tubing size 3/16", length 1.80m, Yankauer handle with control vent and crown tip e. tubing size 3/16", length 3.60m, Yankauer handle with control vent and crown tip	NO	569.56	Rupees Five Hundred Sixty Nine And Paise Fifty Six Only
98	21.00	Laryango Scope with Blades ADULT:Handle: stainless steel or chromium-plated, slightly rigged. b. Depressor and laryngoscope blade: stainless steel, curved or straight. c. Complete, ready-to-use, the laryngoscope should be L-shaped so that the handle and batteries are at a right angle to the blade. d. Illumination: battery compartment in the handle, lamp positioned on the blade. e. An instrument used to inspect the larynx, comprised of a: - Hollow handle, 2-battery compartment type. - Spatula or blade: f. Macintosh type: slightly curved blade with a small bulbous tip, fixed to a hook-on handle. g. Each blade has a single bulb, removable for cleaning.	NO	2847.80	Rupees Two Thousand Eight Hundred Forty Seven And Paise Eighty Only
99	11.00	Laryango Scope with Blades Peadeatric:Handle: stainless steel or chromium-plated, slightly rigged. b. Depressor and laryngoscope blade: stainless steel, curved or straight. c. Complete, ready-to-use, the laryngoscope should be L-shaped so that the handle and batteries are at a right angle to the blade. d. Illumination: battery compartment in the handle, lamp positioned on the blade. e. An instrument used to inspect the larynx, comprised of a: - Hollow handle, 2-battery compartment type. - Spatula or blade: f. Macintosh type: slightly curved blade with a small bulbous tip, fixed to a hook-on handle. g. Each blade has a single bulb, removable for cleaning.	NO	2847.80	Rupees Two Thousand Eight Hundred Forty Seven And Paise Eighty Only
100	20.00	Ambu bag Adult /Peadtric Set (10 +10):700ml and 200ml Self inflating double ended silicone bag with mounts b. Side feed oxygen inlet c. Type "L" non-rebreathing valve with pressure limiting device 40cms of water d. Size 0A Circular Pedi All Silicone facemask e. Reservoir Hose with Adaptor to provide higher oxygen & 1.5mtrs oxygen enrichment tubing f. Carrying case	NO	1423.90	Rupees One Thousand Four Hundred Twenty Three And Paise Ninety Only
101	10.00	New Born Ambu Bag:Stroke volumes: 150 ml b. Patients weight: < 10 kg (1 year) c. Total bag volume: 220 ml d. Oxygen reservoir volume: 100 ml	NO	1423.90	Rupees One Thousand Four Hundred Twenty Three And Paise Ninety Only

Santhosh. D
Contractor

Executive Engineer
Project Division
Yelahanka Zone

	Quantity	Item Name	Unit	Approved Rates of Sri. Santhosh Devarajappa	
				In figures	In words
	10.00	Spot Light	NO	11047.50	Rupees Eleven Thousand Forty Seven And Paise Fifty Only
103	1.00	Haemorroid octomy Set	NO	16694.00	Rupees Sixteen Thousand Six Hundred Ninety Four Only
104	1.00	Appendisectomy set	NO	29460.00	Rupees Twenty Nine Thousand Four Hundred Sixty Only
105	1.00	Shrodhkar Forceps	NO	1669.40	Rupees One Thousand Six Hundred Sixty Nine And Paise Forty Only
106	1.00	Hydrocele SET	NO	8347.00	Rupees Eight Thousand Three Hundred Forty Seven Only
107	1.00	Hernia SET	NO	16595.80	Rupees Sixteen Thousand Five Hundred Ninety Five And Paise Eighty Only
108	1.00	Humbey's Knife	NO	8347.00	Rupees Eight Thousand Three Hundred Forty Seven Only
109	1.00	Sims Anal Speculum Large	NO	1374.80	Rupees One Thousand Three Hundred Seventy Four And Paise Eighty Only
110	1.00	Sims Anal Speculum Medium	NO	1374.80	Rupees One Thousand Three Hundred Seventy Four And Paise Eighty Only
111	1.00	Skin Hooks & retractor	NO	2651.40	Rupees Two Thousand Six Hundred Fifty One And Paise Forty Only
112	1.00	Babcocs	NO	883.80	Rupees Eight Hundred Eighty Three And Paise Eighty Only
113	1.00	Anesthesia Face mask different Sizes	NO	883.80	Rupees Eight Hundred Eighty Three And Paise Eighty Only
114	1.00	A.O.Reduction Clamps Large (ortho)	NO	5008.20	Rupees Five Thousand Eight And Paise Twenty Only
115	1.00	A.O.Reduction Clamps Small (ortho)	NO	5008.20	Rupees Five Thousand Eight And Paise Twenty Only
116	1.00	Intra Medullary Remurus Flexible 8.5,9.5,10.5,11.5 (ortho)	NO	2847.80	Rupees Two Thousand Eight Hundred Forty Seven And Paise Eighty Only

Executive Engineer
Project Division
Yelahanka Zone

Contractor

	Quantity	Item Name	Unit	Approved Rates of Devarajappa	
				In figures	In words
117	1.00	Crocodile Reduction Clamp large	NO	3829.80	Rupees Three Thousand Eight Hundred Twenty Nine And Paise Eighty Only
118	1.00	Crocodile Reduction Clamp small	NO	3829.80	Rupees Three Thousand Eight Hundred Twenty Nine And Paise Eighty Only
119	1.00	wire Cutter	NO	4959.10	Rupees Four Thousand Nine Hundred Fifty Nine And Paise Ten Only
120	1.00	Hand Drill	NO	11096.60	Rupees Eleven Thousand Ninety Six And Paise Sixty Only
121	1.00	Steimion Pins (ORTHO)	NO	1669.40	Rupees One Thousand Six Hundred Sixty Nine And Paise Forty Only
122	1.00	Sinus probe	NO	1325.70	Rupees One Thousand Three Hundred Twenty Five And Paise Seventy Only
123	1.00	Towel Clip	NO	834.70	Rupees Eight Hundred Thirty Four And Paise Seventy Only
124	1.00	S.S. WIRE diffeent Sizes (ORTHO) EACH Size	NO	294.60	Rupees Two Hundred Ninety Four And Paise Sixty Only
125	1.00	K-WIRE DIFFERENT EACH SIZES (ORTHO)	NO	294.60	Rupees Two Hundred Ninety Four And Paise Sixty Only
126	1.00	Lamenectomy SET	NO	27594.20	Rupees Twenty Seven Thousand Five Hundred Ninety Four And Paise Twenty Only
127	1.00	Super micro ear Forceps,serrated, 0.6X2.0MM EFFECTIVE Length 75mm	NO	2651.40	Rupees Two Thousand Six Hundred Fifty One And Paise Forty Only
128	1.00	Ear Forceps, Serrated St 1.0 x 4.2mm Effective Length 75mm	NO	2651.40	Rupees Two Thousand Six Hundred Fifty One And Paise Forty Only
129	1.00	Super micro ear Forceps oval Cupped jars 0.8 x 1.0mm Effective Length75mm	NO	3878.90	Rupees Three Thousand Eight Hundred Seventy Eight And Paise Ninety Only
130	1.00	Super Micro Ear Forceps oval Cupped jars 0.6x 0.8mm Effective Length75mm	NO	3878.90	Rupees Three Thousand Eight Hundred Seventy Eight And Paise Ninety Only

Executive Engineer
Project Division
Yelahanka Zone

Contractor